

BioClot GmbH

Hollerbach 20
D-94501 Aidenbach
www.bioclot.com

Telefon: +49 8543 91171
Fax: +49 8543 91179
bioclot@bioclot.com



Certificate of Analysis

Product: Polysucrose 400 based separating solution
Catalog N°: PS400L
Batch N°: 5580209
Production date: 19.02.2009
Expiry date: 02/2012
Storage: +2°C to RT

QC – Test (method)	Specification	Results
Visual (optical control)	clear, colourless liquid free of particulate matter	pass
pH	6.5 to 7.5	7.14
Osmolality	280 to 310 mOsm/kg	298 mOsm/kg
Density	1.076 to 1.078 g/ml at 20°C	1,077 g/ml
Endotoxin	≤ 0.1 ng/ml	0.026 ng/ml
Sterility	incubation 37°C: negative	pass
	incubation 20°C: negative	pass
MTT	not less than 80%	pass